


ASSAY PERFORMANCE SUMMARY

1,7-Dimethyluric acid



Method	LC/MS/MS
Matrix	human urine
Anti-coagulant	n/a
LLQ	500 ng/mL
ULQ	20,000 ng/mL
Sample Volume	50.0 μ L
Internal Standard	analogue
Accuracy: %Bias	
LLQ	-7.6
LQC	-4.7
MQC	-3.7
HQC	2.7
DilQC (DF=10)	3.2
Precision: %CV	
LLQ	4.9
LQC	3.0
MQC	3.1
HQC	4.8
DilQC (DF=10)	11.7
Stability: Minimum	
Ambient	6 hours
-20°C	In Progress
No. Freeze/thaws	3