


ASSAY PERFORMANCE SUMMARY

Ciprofloxacin



Method	LC/MS/MX
Matrix	human plasma
Anti-coagulant	sodium heparin
LLOQ	10.0 ng/mL
ULQ	1,000 ng/mL
Sample Volume	100 μ L
Accuracy: %Bias	
LLOQ	-8.0
LQC	-10.0
MQC	-1.0
HQC	0.9
DilQC (DF=10)	-11.3
Precision: %CV	
LLOQ	2.7
LQC	8.7
MQC	2.8
HQC	2.0
DilQC (DF=10)	5.5
Stability:	
Ambient	7 hours
No. Freeze/thaws	3