


ASSAY PERFORMANCE SUMMARY

Ibandronate



Method	LC/MS/MS
Matrix	human plasma
Anti-coagulant	K ₂ EDTA
LLQ	0.200 ng/mL
ULQ	200 ng/mL
Sample Volume	200 µL
Internal Standard	Stable label
Accuracy: %Bias	
LLQ	-0.5
LQC	4.7
MQC	-1.9
HQC	2.0
DilQC (DF=10)	2.0
Precision: %CV	
LLQ	7.2
LQC	4.8
MQC	2.0
HQC	1.6
DilQC (DF=10)	1.5
Stability: Minimum	
Ambient	6 hours
-20°C	93 days
No. Freeze/thaws	4