


# ASSAY PERFORMANCE SUMMARY

## Mefenamic Acid



Method	LC/MS/MS
Matrix	human plasma
Anti-coagulant	K <sub>3</sub> EDTA
LLQ	50.0 ng/mL
ULQ	5000 ng/mL
Sample Volume	100 µL
Internal Standard	analogue
Accuracy: %Bias	
LLQ	-3.2
LQC	-4.7
MQC	-2.5
HQC	-5.0
DilQC (DF=10)	-2.4
Precision: %CV	
LLQ	7.5
LQC	5.8
MQC	7.9
HQC	8.2
DilQC (DF=10)	3.4
Stability: Minimum	
Ambient	6 hours
-70°C	119 days
No. Freeze/thaws	4