


# ASSAY PERFORMANCE SUMMARY

## Carboplatin



Method	LC/MS/MS
Matrix	plasma ultrafiltrate
Anti-coagulant	sodium heparin
LLQ	100.0 ng/mL
ULQ	40,000 ng/mL
Sample Volume	50.0 $\mu$ L
Accuracy: %Bias	
LLQ	8.0
LQC	0.7
MQC	-5.6
HQC	-5.3
DilQC (DF=10)	3.8
Precision: %CV	
LLQ	4.8
LQC	11.1
MQC	3.6
HQC	2.0
DilQC (DF=10)	5.2
Stability: Minimum	
Ambient	26 hours
No. Freeze/thaws	4